

2,025 jts 2 7/8"
Baranko Tubing



Baranko Environmental LLC Radiation Survey

Client Location: Baranko Yard	Survey Location: Baranko Yard
Meter Make & Model#: Ludlum Model 2241-2	Meter Serial #: 311200
Probe Model & Serial #: 44-2 PR300918	Meter Calibration Date: 03/01/2023
Check Source Serial #: RS-188	Assigned Reading: 105µR/hr
Pre-Survey Battery Power Checked: X	Pre-Survey Measured Reading: 96µR/hr
Pre-Survey Battery Power Strength: X	Background: 20µR/hr

18	18	16	16	16	16	17	16
18	18	18	19	18	18	18	18
18	18	18	19	18	18	18	18
17	16	17	18	18	18	18	16
18	16	18	16	18	16	17	18
16	16	18	16	18	16	17	18

Not to Scale

I hereby certify under penalty of law that I personally examined and am familiar with the information submitted in this document and all attachments and that, based on my personal inquiry of those individuals immediately responsible for obtaining the information, I hereby represent in writing that the information contained within is true, accurate, and complete to the best of my knowledge, information, and belief. I am aware that there are significant penalties for submitting false information, including the possibility of fines and imprisonment, or both.

Post-Survey Battery Power Checked: X	Assigned Reading: 105µR/hr
Post-Survey Battery Power Strength: X	Post-Survey Measured Reading: 95µR/hr

Print Name: Joe Robinson	Title: Inventory Manager/RSO
Signature: <i>Joseph Robinson</i>	Date & Time of Survey: 04/26/23

CRSO REVIEW: _____
Signature

DATE: _____